

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/11/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:56
INTERNAL STANDARD	VERIFIED	01:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:57
BLANK TEST	0.00	01:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:58
BLANK TEST	0.00	01:59
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	02:00



SGT Ralph J Bartlett WESO SHERIFF #162 08-11-11