

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/22/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:06
INTERNAL STANDARD	VERIFIED	04:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:07
BLANK TEST	0.00	04:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:08
BLANK TEST	0.00	04:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:09
BLANK TEST	0.00	04:10

