

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/29/2011

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:39
INTERNAL STANDARD	VERIFIED	17:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:40
BLANK TEST	0.00	17:40
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	17:41
BLANK TEST	0.00	17:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:42
BLANK TEST	0.00	17:43

JM Lopez
#299
8/29/11

