

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/08/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:13
INTERNAL STANDARD	VERIFIED	03:13
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:13
BLANK TEST	0.00	03:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:14
BLANK TEST	0.00	03:15
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:15
BLANK TEST	0.00	03:16



SOS Ralph Bartlett

WCJO #162 09-08-11