

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/27/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:02
INTERNAL STANDARD	VERIFIED	02:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:02
BLANK TEST	0.00	02:03
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:03
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:04
BLANK TEST	0.00	02:05



SI - Ralph J Bartlett

WESO #162 09-27-11