

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/02/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:21
INTERNAL STANDARD	VERIFIED	02:21
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:22
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:24
BLANK TEST	0.00	02:25



on R. J. Bartlett was checked #162 10-2-11