

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/10/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:23
INTERNAL STANDARD	VERIFIED	22:23
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:23
BLANK TEST	0.00	22:24
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:24
BLANK TEST	0.00	22:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:26
BLANK TEST	0.00	22:26



Handwritten notes: 10/10/11 #102