

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/25/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:20
INTERNAL STANDARD	VERIFIED	23:20
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:20
BLANK TEST	0.00	23:21
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	23:21
BLANK TEST	0.00	23:22
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:22
BLANK TEST	0.00	23:23



R. J. Bartlett *10/25/2011* *10/25/2011* *144506*