

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 10/31/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:19
INTERNAL STANDARD	VERIFIED	03:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:19
BLANK TEST	0.00	03:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:20
BLANK TEST	0.00	03:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:21
BLANK TEST	0.00	03:22



SG5 Rugh / Ruoco

WCSO #162 10/31/11