

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 12/02/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:26
INTERNAL STANDARD	VERIFIED	03:26
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:26
BLANK TEST	0.00	03:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:27
BLANK TEST	0.00	03:28
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:28
BLANK TEST	0.00	03:29



SCA Ralph Bartlett WES #162 12-02-11