

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 12/25/2011

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:48
INTERNAL STANDARD	VERIFIED	21:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:48
BLANK TEST	0.00	21:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:49
BLANK TEST	0.00	21:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:50
BLANK TEST	0.00	21:51



505 *Russ / B...* WCSO #102 12-25-11