

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 12/30/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:33
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:35



SGS Ruff J Bartlett WCSO SHERIFF #102 12-30-11