

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 01/07/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:23
INTERNAL STANDARD	VERIFIED	00:23
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:24
BLANK TEST	0.00	00:24
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:25
BLANK TEST	0.00	00:25
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:26
BLANK TEST	0.00	00:26



Supervisor: Bartlett *WPCO #11190* *01-07-12*