

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/21/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN COUNTY SHERIFF'S OFFICE
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:03
INTERNAL STANDARD	VERIFIED	00:03
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:04
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:05
BLANK TEST	0.00	00:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:06
BLANK TEST	0.00	00:06



565 Ruger 1 Barrett WCSO SHERIFF #102 01-21-12