

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/09/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:33
INTERNAL STANDARD	VERIFIED	00:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:34
BLANK TEST	0.00	00:35
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:35
BLANK TEST	0.00	00:36

