

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/15/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:12
INTERNAL STANDARD	VERIFIED	00:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:13
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:14
BLANK TEST	0.00	00:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:15
BLANK TEST	0.00	00:16

