

SOLUTION CHANGE




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 105208  
 -----

Date: 02/29/2012

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:14
INTERNAL STANDARD	VERIFIED	01:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:15
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:17



10520 # 162