

# SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/08/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11240

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 00:59 |
| INTERNAL STANDARD       | VERIFIED | 01:00 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 01:00 |
| BLANK TEST              | 0.00     | 01:01 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.09     | 01:01 |
| BLANK TEST              | 0.00     | 01:02 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 01:02 |
| BLANK TEST              | 0.00     | 01:03 |

