

**SUPERVISOR TEST**




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/22/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:21
INTERNAL STANDARD	VERIFIED	03:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:21
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:22
BLANK TEST	0.00	03:23
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:23
BLANK TEST	0.00	03:24



S67 *Ralph Bartlett* WOSO #162 03-22-12