

SUPERVISOR TEST




-----  
WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208  
-----

Date: 03/28/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:40
INTERNAL STANDARD	VERIFIED	01:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:41
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:43
BLANK TEST	0.00	01:44



SG Bartlett

WESO #162

03-28-12