

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/14/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:42
INTERNAL STANDARD	VERIFIED	00:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:44
BLANK TEST	0.00	00:45

