

**SUPERVISOR TEST**




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 06/06/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:15
INTERNAL STANDARD	VERIFIED	01:15
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:17
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:18



*SG - Ralph Bartlett* WCSO SHELD # 162 06-06-12