

**SUPERVISOR TEST**




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 06/20/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:35
INTERNAL STANDARD	VERIFIED	00:35
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:36
BLANK TEST	0.00	00:36
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:37
BLANK TEST	0.00	00:37
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:38
BLANK TEST	0.00	00:38



SGT Ralph J Bartlett WESO SHEAR #162 06-20-12