

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 07/03/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:28
INTERNAL STANDARD	VERIFIED	01:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:29
BLANK TEST	0.00	01:30
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:30
BLANK TEST	0.00	01:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:31
BLANK TEST	0.00	01:32



SG5 *Ralph Bartlett* Weso #162 07-03-12