

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 07/09/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:17
INTERNAL STANDARD	VERIFIED	03:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:17
BLANK TEST	0.00	03:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:18
BLANK TEST	0.00	03:19
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:19
BLANK TEST	0.00	03:20



S65 *Ralph Bartlett* WESO #162 07-09-12