

SUPERVISOR TEST

-----  
WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208  
-----  
Date: 07/25/2012

Operator Name: JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:25
INTERNAL STANDARD	VERIFIED	22:26
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.09	22:26
BLANK TEST	0.00	22:27
REFERENCE STANDARD	33.95C	
REFERENCE STANDARD	0.09	22:27
BLANK TEST	0.00	22:28
REFERENCE STANDARD	33.97C	
REFERENCE STANDARD	0.09	22:28
BLANK TEST	0.00	22:29

*J.M. Lopez* #299

