

SUPERVISOR TEST - *Solution Citric Acid* **(11)**




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 07/31/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:27
INTERNAL STANDARD	VERIFIED	22:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:27
BLANK TEST	0.00	22:28
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:28
BLANK TEST	0.00	22:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:29
BLANK TEST	0.00	22:30



*SGT RALPH J BARTLETT WESO #162 07-31-12*