

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/09/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:14
INTERNAL STANDARD	VERIFIED	00:15
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:15
BLANK TEST	0.00	00:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:16
BLANK TEST	0.00	00:17
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:17
BLANK TEST	0.00	00:18



Sgt Ralph Bartlett WCSO #162 08-09-12