

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/22/2012

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:23
INTERNAL STANDARD	VERIFIED	18:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:23
BLANK TEST	0.00	18:24
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	18:24
BLANK TEST	0.00	18:25
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	18:25
BLANK TEST	0.00	18:26

JML 299

