

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/29/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:37
INTERNAL STANDARD	VERIFIED	00:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:38
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41



SGS Ralph J Bartlett WES0 #162 08-29-12