

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 09/06/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:17
INTERNAL STANDARD	VERIFIED	00:17
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:19
BLANK TEST	0.00	00:20



SGS Dept of Justice weso #162 09-06-12