

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/17/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:28
INTERNAL STANDARD	VERIFIED	01:28
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:28
BLANK TEST	0.00	01:29
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:29
BLANK TEST	0.00	01:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:30
BLANK TEST	0.00	01:31



SGS *Ralph Bartlett* WESO #162 ~~09-17-12~~ 09-17-12