

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 09/24/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:16
INTERNAL STANDARD	VERIFIED	23:16
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:16
BLANK TEST	0.00	23:17
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:18
BLANK TEST	0.00	23:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:19
BLANK TEST	0.00	23:19



SGT Ralph J Bartlett WCSU #162 09-24-12