

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/01/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:38
INTERNAL STANDARD	VERIFIED	00:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:41



SGT Ruff / T. B. W. WC10 #162 10-01-12