

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/09/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:16
INTERNAL STANDARD	VERIFIED	02:16
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:16
BLANK TEST	0.00	02:17
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:17
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:19



SGT *Ralph J Bartlett* WCSO SHERIFF #102 10-09-12