

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/23/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:39
INTERNAL STANDARD	VERIFIED	02:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42



SGT Ralph J Bartlett Weso #162 10-23-12