

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/29/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:14
INTERNAL STANDARD	VERIFIED	01:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:15
BLANK TEST	0.00	01:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:17



SGT *Ralph J Bartlett* WCSO #162 10-29-12