

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/01/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:50
INTERNAL STANDARD	VERIFIED	00:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:53
BLANK TEST	0.00	00:54



SG5 Ralph J Bartlett WCSO #142 11-01-12