

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/21/2012

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:32
INTERNAL STANDARD	VERIFIED	09:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	09:32
BLANK TEST	0.00	09:33
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	09:33
BLANK TEST	0.00	09:34
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	09:35
BLANK TEST	0.00	09:35

Charles B. Geisler #298

