

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/27/2012

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 08:48 |
| INTERNAL STANDARD | VERIFIED | 08:48 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 08:48 |
| BLANK TEST | 0.00 | 08:49 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 08:49 |
| BLANK TEST | 0.00 | 08:50 |
| REFERENCE STANDARD TEMP | 33.96c | |
| REFERENCE STANDARD | 0.09 | 08:50 |
| BLANK TEST | 0.00 | 08:51 |

Charles B. Geisler #210

