

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/13/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:20
INTERNAL STANDARD	VERIFIED	00:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:21
BLANK TEST	0.00	00:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:22
BLANK TEST	0.00	00:23
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:23
BLANK TEST	0.00	00:24



SGT Ralph J Bartlett

wcso #162 12-13-12