

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/26/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:39
INTERNAL STANDARD	VERIFIED	22:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:39
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:40
BLANK TEST	0.00	22:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:41
BLANK TEST	0.00	22:42



505 Rpt 7 Results w/30 #162 12-26-12