

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/23/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:30
INTERNAL STANDARD	VERIFIED	00:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:33

