

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 01/07/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:17
INTERNAL STANDARD	VERIFIED	01:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:21



SGT Ralph J Bartlett WESD #162 01-07-12