

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/13/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:49
INTERNAL STANDARD	VERIFIED	00:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:52



565 *Ralph Bartlett* WCSO #162 01-13-12