

SOLUTION CHANGE

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/01/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:17
INTERNAL STANDARD	VERIFIED	01:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:21

