

SUPERVISOR TEST

 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/15/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:35
INTERNAL STANDARD	VERIFIED	00:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:35
BLANK TEST	0.00	00:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:36
BLANK TEST	0.00	00:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:37
BLANK TEST	0.00	00:38

