

**SUPERVISOR TEST**




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 03/22/2012

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:38
INTERNAL STANDARD	VERIFIED	03:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:39
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:40
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:40
BLANK TEST	0.00	03:41



SGT *Ralph Bartlett* WESO #162 03-22-12