

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/28/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:58
INTERNAL STANDARD	VERIFIED	01:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:58
BLANK TEST	0.00	01:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	02:00
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:01



*SGT Ralph Bartlett WOSO #162 03-28-12*