

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/16/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:46
INTERNAL STANDARD	VERIFIED	00:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:46
BLANK TEST	0.00	00:47
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:47
BLANK TEST	0.00	00:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:48
BLANK TEST	0.00	00:49

